

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

***Signature® Flowmeter with TIENet™ 330 Bubbler Module***

Manufactured by:

***Teledyne ISCO***

4700 Superior Street,  
Lincoln,  
Nebraska,  
68504-1398  
USA

has been assessed by CSA Group  
and for the conditions stated on this certificate complies with:

**Performance Standards and Test Procedures for Continuous Water  
Monitoring Equipment, Part 3: Performance standards and test procedures for water  
flowmeters, Environment Agency, version 4, March 2020**

The combined performance characteristic ( $U_c$ , the expanded uncertainty) is **0.36%** (Class2)

Certification range:

0.003 to 3 m

Project No.: 80215329  
Certificate No: CSA MC140251/03  
Initial Certification: 11 August 2014  
This Certificate issued: 16 August 2024  
Renewal Date: 10 August 2029



Andrew Young  
Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

**CSA Group Testing UK Ltd**

Unit 6, Hawarden Industrial Park  
Hawarden, Deeside, CH5 3US  
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.  
For conditions of use, please consider all the information within.  
This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Certificate Contents

Approved Site Application.....	2
Basis of Certification .....	2
Product Certified.....	3
Certified Performance .....	4
Description.....	6
General Notes .....	7

## Approved Site Application

Any potential user should make sure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency guidance available at [www.mcerts.net](http://www.mcerts.net)

The product is suitable for use, where it is appropriate, for regulated applications such as abstraction, effluent discharge, ultraviolet disinfection and industrial processing.

## Basis of Certification

This certification is based on the following test report(s) and on CSA Group's assessment and ongoing surveillance of the product and the manufacturing process:

- R1 WRc Report Number UC9578 v03 dated July 2013
- R2 WRc Report Number UC9663 v03 dated July 2013
- R3 Signature Flow Meter Field Test Report 20131022 dated 21 October 2013
- R4 Testing of Response Time for Signature Bubbler

Certificate No: CSA MC140251/03  
Certificate issued: 16 August 2024

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Product Certified

The Signature® Flowmeter with TIENet™ 330 Bubbler module system consists of the following parts:

- Signature® flowmeter (100 to 230 V AC). Hardware version A0.
- TIENet™ model 330 bubbler type level sensor
- Optional external power loss alarm (required for MCERTS conformity).

This certificate applies to all instruments fitted with software version 1.18.009 onwards (Signature® flowmeter serial number 212C02317 & TIENet™ model 330 bubbler type level sensor serial number 212D01154 onwards).

Certificate No: CSA MC140251/03  
Certificate issued: 16 August 2024

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: -20°C to +60°C

Instrument IP rating: NEMA 4X/IP66

**The instrument meets MCERTS Class 2 requirements for the combined performance characteristic as specified in Table 7 of the MCERTS performance standard. Details of individual performance characteristics are summarised below:**

Results are expressed as error % certification range 0.003 to 3m, unless otherwise stated.

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Protection against unauthorised access	Password required					Clause 3.1.2
Units of measurement	The flowmeter records in metric units					Clause 3.1.6
Indicating device	The flowmeter displays totalised volume and/or flow-rate					Clause 3.1.3
<b>Combined performance characteristic</b>					<b>Note 1 0.36</b>	<b>Table 7 Class 2</b>
Mean error	0.12					Clause 6.3.2 Class 2
Repeatability	0.05					Clause 6.3.2 Class 1
Supply voltage 100 to 240 V AC	0.02					Clause 6.3.3 Class 1
Output impedance	<0.01				<b>Note 2</b>	Clause 6.3.4 Class 1
Ambient temperature	0.18					Clause 6.3.6 Class 3
Relative humidity	0.12					Clause 6.3.6 Class 3
Accuracy of computation	-0.20				<b>Note 2</b>	Clause 6.3.11 Class 3
User defined stage-discharge equation	0.15				<b>Note 2</b>	Clause 6.3.12 Class 3

Certificate No: CSA MC140251/03  
Certificate issued: 16 August 2024

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time					Note 1 0.36	Clause 6.3.19 <30 seconds
Error under field test conditions	Max error            0.63% Min error            0.03% Mean error           0.17% Proportion of errors $\leq 0.2\%$ = 58% Proportion of errors $\leq 0.5$ = 95%					Clause 7.3 0.5% Class 2
Up time	0.05				100%	Clause 7.4 >95%
Maintenance	30 mins scheduled during field trial					Clause 7.5

Note 1: The combined performance characteristic reported is the root-sum-square addition of the maximum errors recorded in the following tests: mean error, repeatability, supply voltage, output impedance, ambient air temperature, computational accuracy and resolution.

Note 2: A simulated level input using a Teledyne ISCO TIENet™ 310 ultrasonic level sensor was used to evaluate the software / electronics response of the Signature unit.

Certificate No: CSA MC140251/03  
 Certificate issued: 16 August 2024

*This certificate may only be reproduced in its entirety and without change  
 To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Description

The Signature® Flowmeter is designed for open channel flow monitoring applications using any combination of flow and parameter measurement technologies and sampling, depending on what is required at the measurement site.

The bubbler level sensor can be used in almost any open channel flow application and performance is unaffected by steam or foam on the surface of the water. The bubble line is anchored in the flow stream at the appropriate measuring point in the weir, flume, or other open channel flow situation. The pressure in the bubble line is proportional to the liquid level in the flow stream sensing the liquid level.

The Signature Flowmeter can calculate flow using standard open channel level-to-flow conversions, as well as user-defined equations, level to area data points or level to flow data points, depending upon the measurement device(s) used in the application and the program specified by the user.

The Signature Flowmeter allows multiple simultaneous flow technologies, input for pH and temperature, accepts input from SDI-12 devices and Modbus devices, and can support multiple analog output signals. The Signature has a graphical display for viewing of parameter measurements and instrument configuration. It is compatible with Teledyne Isco Flowlink software which allows on site or remote data collection or instrument configuration. The Signature has a front panel notification LED to indicate and alarm condition. The Signature is capable of communicating through an Ethernet modem, a cellular modem, or an USB interface.

The Signature Flowmeter has features to verify data integrity. It logs events such as changes in calibration and power outages to validate data accuracy. Data can be reviewed to detect any type of data alteration. Program reports, summary reports, and time series data are retrievable using a USB flash drive.

With multiple smart interface options and multi-parameter logging (such as pH), the Signature Flow Meter provides a common platform for control action, reporting, and communication.

## General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of CSA Certificates'.
2. The design of the product certified is defined in the CSA design schedule for certificate No. CSA MC140251/03.
3. If the certified product is found not to comply, CSA Group should be notified immediately at the address shown on this certificate.
4. The certification marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of CSA Certificates'.
5. This document remains the property of CSA Group and shall be returned when requested by CSA Group.

Certificate No: CSA MC140251/03  
Certificate issued: 16 August 2024

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*