

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

ISCO 5800 Automatic Waste Water Sampler

Manufactured by:

Teledyne ISCO Inc.

4700 Superior Street
Lincoln
Nebraska
NE 68504
USA

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Water
Monitoring Equipment – Part 1, Version 4 dated April 2017
EN16479:2014**

Certification Ranges :

Lift height 0 to 6 meters

Project No. : 16W29523 / 70171503
Certificate No : Sira MC130227/02
Initial Certification : 10 July 2013
This Certificate issued : 22 March 2019
Renewal Date : 09 July 2023

Emily Alexander
Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certificate Contents

Approved Site Application.....	2
Basis of Certification	2
Product Certified.....	2
Certified Performance	3
Description.....	5
General Notes	5

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

The product is suitable for use on applications for compliance with the Urban Wastewater Treatment Regulations.

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

WRc report	Report Reference: UC 7399 dated May 2007
WRc report	Report Reference: UC 9502 dated March 2013
WRc report	Report Number UC 13500 dated November 2018
CSA Sira	Evaluation report reference: 70171503 dated February 2019

Product Certified

The 5800 automatic wastewater sampler consists of the following parts:

- 5800 control panel
- 5800 pump assembly
- 5800 distributor assembly
- 5800 refrigeration system

This certificate applies to all instruments fitted with software version 1.00.0006 onwards (serial number 213A00001 onwards).

Certificate No : Sira MC130227/02
This Certificate issued : 22 March 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: -10°C to +40°C

Test	Results	MCERTS specification
Sample Collection	Flow proportional and timed sampling available Certified for use with single 10 litre composite sample bottle and 24 X 1 litre bottles.	Clause 3.1.2
Sample interval	Sample interval range is 1 min to 99 hours, 59 minutes with increments of 1 min is selectable 4-20mA and pulse outputs are available 1min to 9,999 flow pulses with increments of 1 pulse.	Clause 3.1.2
Sample failure	Sample failures are recorded. Fault indicated on display	Clause 3.1.2
Sample line diameter	9.52 mm.	Clause 3.1.2 >9mm
Sample volume	Sample volume adjustable over the range 10 to 9,990ml in 1ml increments	Clause 3.1.2
Maximum volume of a discrete sample that can be set Total storage capacity both by numbers and volumes of individual bottles and in a composite container	9,990ml The following are available: 24 X 1 L PP or 350ml glass 4 X 10 L PE or glass 2 X 10 L PE or glass 1 X 20 L PE or glass 1 X 10 L PE or glass 24 X ProPak, 1 L wedge 1 X ProPak, 10 L round	Clause 3.1.2
Maximum sampling head	6 metres	Clause 3.1.2
Sample volume error – C.T.C.V Time proportional	Expanded uncertainty: Mean error: 2.03% at 1m 2.69% at 1m 3.29% at 3.5m -2.46% at 3.5m 4.11% at 7m -4.37% at 7m Overall: 3.14% Overall: -1.38%	Clause 6.4.1.1 <5%

Certificate No : Sira MC130227/02
This Certificate issued : 22 March 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results		MCERTS specification
Sample volume – Contact time variable volume (CTVV) flow proportional sampling	Expanded uncertainty: 2.46% at 1m 1.69% at 3.5m 1.31% at 7m Overall: 1.83%	Mean error: -0.03% at 1m -1.05% at 3.5m 0.71% at 7m Overall: -0.12%	Clause 6.4.1.2 <5%
Sample volume – Constant volume variable time (CVVT) flow proportional sampling	Expanded uncertainty: 1.77% at 1m 2.41% at 3.5m 2.26% at 7m Overall: 2.15%	Mean error: -0.04% at 1m -3.89% at 3.5m -2.34% at 7m Overall: -2.09%	Clause 6.4.1.3 <5%
Analogue flow signal – Constant volume variable time (CVVT) flow proportional sampling	0.03%		Clause 6.4.2.4 <1%
Event Triggered Sampling	0.03%		Clause 6.4.2.5 <1%
Sample line velocity	0.95 m/s at 1m sampling head 0.94 m/s at 2m sampling head 0.90 m/s at 3m sampling head 0.88 m/s at 4m sampling head 0.85 m/s at 5m sampling head 0.51 m/s at 6m sampling head		Clause 6.4.3 >0.5 m/s
Power Supply 103.5 VAC to 126.5 VAC	3/8-inch 6m suction line Min: 0.564m/s Ref: 0.57m/s Max: 0.58m/s		Clause 6.4.4.1 >0.5m/s
Sample integrity	No statistically significant difference was found in analysis for BOD, COD, suspended solids, total N and total P		Clause 6.4.5 Annex B5
Sample timing	4 seconds		Clause 6.4.6 < ±10 sec/24h
Ambient temperature effects a) volume	Expanded uncertainty: 2.43% at -10°C 2.58% at 40°C	Mean error: 1.43% at -10°C 3.04% at 40°C	Clause 6.4.7.2 <5% <5%
Ambient temperature effects b) temperature	During sample period: 4.1°C at -10°C 4.3°C at 20°C 4.0°C at 40°C	24hrs after sample period: 3.0°C at -10°C 3.8°C at 20°C 3.7°C at 40°C	Clause 6.4.7.3 Maintain sample between 0°C to +5°C

Certificate No : Sira MC130227/02
This Certificate issued : 22 March 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The 5800 Refrigerated Sampler System consists of a 5800 Control Panel, 5800 Pump Assembly, 5800 Distributor Assembly, and 5800 Refrigeration System.

The 5800 Control Panel provides user programmable sampler features via a keypad and display. The 5800 Pump Assembly consists of a peristaltic pump for pulling samples to the sampler. The 5800 Distributor Assembly distributes the pumped sample into container(s) located in the refrigerator. The sample is transported by 3/8 inch I.D. PVC or PTFE lined suction line.

The refrigeration system utilizes a 1/4 hp compressor running on R134a refrigerant and is controlled by the microprocessor located in the 5800 Sampler Control Panel. The refrigerator cabinet is constructed from UV resistant LLDPE. The refrigeration system constructed from stainless steel and is powered by 230Vac 50 Hz.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule V00 for certificate No. Sira MC130227/00
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC130227/02
This Certificate issued : 22 March 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*